
Formaldehyd Probenpuffer

Gefahrstoff



Separation of Total RNA on the FlashGel System for RNA – Samples of RNA Marker (Lonza) (lanes 1 and 2) and E. coli total RNA (lanes 3 and 4) contain 50 ng (lanes 1 and 3) or 250 ng (lanes 2 and 4) of RNA per 5 µl load. Samples prepared with Formaldehyde Sample Buffer (Lonza) and denatured 5 minutes at 65°C. FlashGel RNA Cassette run for 8 minutes, followed by a 20 minute hold prior to imaging.

Lonza

5 x 1 ml

Artikel-Nr.: 850571 | Lonza | Hersteller-Nr.: 50571

189,00 €*

*zzgl. MwSt. zzgl. Versandkosten

Beschreibung

Eigenschaften: für FlashGel RNA Marker
Kategorie: FlashGel System
Produkttyp: RNA Probenpuffer
Verpackung: 5 x 1 ml

Kurzanleitung Formaldehyd Probenpuffer

1. RNA-Proben auf Eis auftauen
2. Probenpuffer bei Raumtemperatur auftauen
3. Probenpuffer gründlich vortexen
4. RNA-Probe 1:1 mit Formaldehyd-Probenpuffer gründlich mischen
5. 5–15 Minuten auf 65 °C erhitzen
6. 1 Minute auf Eis abkühlen
7. Proben auf das Gel auftragen

Lagerung: -20 °C
Nur für Forschungszwecke

Telefon: +49 (0) 51 52 / 90 20
Telefax: +49 (0) 51 52 / 20 70
E-Mail: support@biozym.com
Internet: www.biozym.com

Rechtliches
Registergericht: Amtsgericht Hannover
Registernummer: HRB 101682
UST-ID: DE 813739502

Biozym Scientific GmbH
Geschäftsführer
Dr. Sebastian Petri